

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME

## CB TEST CERTIFICATE

Product

Switching Power Supply

Name and address of the applicant

MEAN WELL Enterprises Co., Ltd.  
No.28, Wuquan 3rd Rd., Wugu District, New Taipei City 24891,  
Taiwan

Name and address of the manufacturer

MEAN WELL Enterprises Co., Ltd.  
No.28, Wuquan 3rd Rd., Wugu District, New Taipei City 24891,  
Taiwan

Name and address of the factory

*Note: When more than one factory, please report on page 2*MEAN WELL Enterprises Co., Ltd.  
No.28, Wuquan 3rd Rd., Wugu District, New Taipei City 24891,  
Taiwan [Additional Information on page 2](#)

Ratings and principal characteristics

Input: 100-240 Vac, 50/60 Hz, 1.2 A for MPM-45-yz series; 1.5 A  
for MPM-65-yz series; 1.9 A for MPM-90-yz series  
Output : See Test Report for details.

Trademark / Brand (if any)



or MEAN WELL

Type of Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

MPM-x-yz  
See Page 2

Additional information (if necessary may also be reported on page 2)

Additionally evaluated to EN 60601-1:2006 / A1:2013 / A12:2014.  
The risk management requirements of the standard were not  
addressed. [Additional Information on page 2](#)

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005

As shown in the Test Report Ref. No. which forms part of this Certificate

191103401 issued on 2020-01-03

This CB Test Certificate is issued by the National Certification Body



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2020-01-09

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

**DK-92007-UL**

**Model Details:**

MPM-x-yz (Where x can be 45, 65 or 90; y can be 5, 12, 15, 24 or 48; z can be ST or Blank)

**Factories:**

MEAN WELL (GUANGZHOU) ELECTRONICS CO., LTD HUADU BRANCH.

No.11 Jingu South Road, Huadong Town, Huadu District, Guangzhou Guangdong 510890, China

SuZhou MEAN WELL Technology Co., Ltd.

No. 77, Jian-min Road, Dong-qiao, Pan-yang Ind. Park, Huang-dai Town, Xiang-cheng District, Suzhou, Jiangsu 215152, China

YONGDEN TECHNOLOGY CORPORATION

345 MacArthur Highway, Tabang, Guiguinto, Bulacan 3015, Philippines

**Additional information (if necessary)**



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2020-01-09

Signature:

Jan-Erik Storgaard